## Product list 2024 CONFORM<sup>®</sup> Med. Air Cushion System Mattress Overlay System



The system is used as an overlay on top of the standard bed mattress. A complete mattress system consists of 4 sections which are held together by connectorstrips. With the four individual segments, there is the option of choosing a foam variant in the head and foot areas, which then do not need to be filled with air. The variants are listed below. The inflatable segments are equipped with two valves to enable inflation/deflation from either side of the bed. The use of a bed sheet on top of the mattress is mandatory. In case of the cremes on the patients skin, a waterproof cover must be used to avoid contact of creme with the natural rubber which could lead to material damage.

	Article No.	Code	Details
	11503	BMC 2	195 x 85 cm 2 inflatable mattress sections 2 foam sections with waterproof cover connector strips and pump
	11504	BMC 3	195 x 85 cm 3 inflatable mattress sections 1 foam section with waterproof cover connector strips and pump
BMC 3	11505	BMC 4	195 x 85 cm 4 inflatable mattress sections connector strips and pump

CONFORM<sup>®</sup> Medical Air Cushion Systems are Medical Products - Class 1 (Reg. No. DE/CA /23/051603NA001.01; No. 0037437)

## Accessories

Article No.	Description
14756	CONFORM <sup>®</sup> XXL Box, Height, 7.5 cm
10572	CONFORM <sup>®</sup> Pump
10623	Foam section - Height 7,5 cm (85 x 48 cm - Foam made of Polyurethan)
2405	CONFORM <sup>®</sup> single mattress section - Height 7,5 cm (width 48 x depth 85 cm)
10576	Mattress Incontinence Cover
	(100x 200 cm, Fematex waterproof, vapur permeable 2 way stretch cover with nylon backing)
10579	Connector strips



Wolff Kunststoffe GmbH Industriestraße 25

69509 Mörlenbach/Germany

The company is certified according to:





## Phone0049 (0Emailinfo@vHomepagewww.wSRNDE-MF-

0049 (0)6209-72000 info@wolff-kunststoffe.de www.wolff-kunststoffe.de DE-MF-000178949



CE-Declaration: The manufacturer hereby declares, that CONFORM<sup>®</sup> Med. Air Cushion Systems are in accordance with all requirements for medical products class 1 as set out in the guideline 2017/745/EU -Appendix VIII, Chapter III, Paragraph 4.1 - Regulation 1.